



**GOVERNMENT OF INDIA
CENTRAL WATER COMMISSION
MEGHNA INVESTIGATION DIVISION
YINGYAR VILLA, LUMPYNGGAD
SHILLONG-793014**

No. MID/JEHQ-1/2011-12/2500-04

Dated the 27th December, 2011

NOTICE INVITING QUOTATION

Sealed quotations are hereby invited by the undersigned on behalf of the President of India for supply of following items as detailed below. **The quotations should reach in the office of the undersigned on or before 27-01-2012 up to 15.00 hrs** and will be opened on the same day i.e. **27-01-2012 at 15.30 hrs** in the presence of quotationers or their authorized representatives, if any.

The undersigned reserves the right to accept or reject any or all the quotations without assigning any reason thereof.

SN	Particulars of item	Qty.	Unit	Rate (Rs.)
1.	<p><u>Counter Type Anemometer</u></p> <p><u>Specifications</u></p> <p>1. Sensing range 50 m/s accuracy up to 2.5 m/s, 0.25 m/s above 2.5 m/s, 10% FS</p> <p>2. Materials cups copper or aluminium sheet 0.71 mm thickness centre spindle stainless steel rod worm on spindle manganese bronze alloy worm wheel hard brass</p> <p><u>Conditions & Requirements</u></p> <ul style="list-style-type: none"> • The instrument shall be of such a design that it operates reliably and accurately under the prevailing environmental and weather conditions. • The instrument shall be easy to operate and maintain. • The instrument shall be supplied with the accessories as needed for effective deployment. • All materials on the instrument exterior shall be non-corrosive. • The instrument shall have an expected technical lifetime of not less than 10 years. • The instrument shall be capable to operate for at least 3 months without any servicing. • An operator's manual, related to the type and model of the instrument, shall be part of the delivery. • A suitable non-reset direct-drive revolution counter 	6nos.		

	<p>adding one count for each one-tenth revolution of the drive shaft in one direction shall be used.</p> <ul style="list-style-type: none"> • The ratio between gears shall be such that 1 count equals 100 m/s. • The anemometer shall comply to IS 5912-1970. 			
2	<p><u>STEVENSON SCREEN</u></p> <p><u>Specifications</u></p> <p>1. Standard width length between inner slats of louvers 560 mm width between inner slats of louvers 315 mm height between middle of lower roof and floor 412 mm</p> <p><u>Conditions & Requirements</u></p> <ul style="list-style-type: none"> • Double layered roof with ventilation provision. • The screen shall comply with IS 5948 – 1970. • All dimension requirements shall be met within 5% of specified dimension. 	2 nos.		
3	<p><u>WIND VANE</u></p> <p><u>Specifications</u></p> <p>pointer counter balanced direction arms 8, with direction letters for N, E, S, W mounting materials aluminium (ring for direction arm, arm support & horizontal arm) fin & bearing housing cap aluminium alloy spindle brass direction letters galvanised steel painted red</p> <p><u>Conditions & Requirements</u></p> <ul style="list-style-type: none"> •The instrument shall be of such a design that it operates reliably and accurately under the prevailing environmental and weather conditions. •The instrument shall be easy to operate and maintain. •The instrument shall be supplied with the accessories as needed for effective deployment. •All materials on the instrument exterior shall be non-corrosive. •The instrument shall have an expected technical lifetime of not less than 10 years. •The instrument shall be capable to operate for at least 6 months without any servicing. •An operator's manual, related to the type and model of the instrument, shall be part of the delivery. •When the complete instrument on final assembly is held firm at bottom and the balance weight imparted an impact by a hammer with a mass of 0.72 kg and having a handle length of 240 mm, it shall make not less than five and not more than ten complete revolutions. • The general arrangement and main dimensions of wind vane shall comply with IS 5799-1970. 	6 nos.		

Terms & Conditions:

1. The quotations duly signed should be properly sealed and either dropped in quotation Box in the office of the Executive Engineer, Meghna Investigation Division, Yingyar Villa, Lumpynggad, Shillong-14 or send by post. The sealed envelope should be superscribed "**Quotation for Meteorological Instruments due on 27-01-2012**".
2. In the event of the date of opening of quotation being a holiday, quotations will be opened on the next working day with timings unchanged.
3. Quotationers must have PAN/TIN/Service Tax number/VAT.
4. Photocopies of Tax Clearance Certificate and Authorized Dealership Certificate from manufacturer should be enclosed along with the quotations.
5. The pamphlets / brochures should be submitted along with the quotation.
6. The undersigned does not bind himself to accept the lowest quotation and reserve the right to accept or reject any or all the quotations without assigning any reason thereof.
7. The material must be IMD certified.
8. The material must be of standard quality and as per specifications and free from all defects. The material if found defective subsequently or not conforming to the specifications as stated above, have to be replaced by the firm without any additional charges.
9. The material shall have a warranty of minimum one year from the date of supply.
10. Material should be delivered at office of the Executive Engineer, Meghna Investigation Division, Yingyar Villa, Lumpynggad, Shillong-14 only.
11. The quantity of materials may be increased or decreased at the time of supply order.
12. Material should be supplied within 30 days from the date of supply order.
13. The rates should be F.O.R. destination at Office of Executive Engineer, Meghna Investigation Division, Yingyar Villa, Lumpynggad, Shillong-14 and inclusive of all taxes/charges including VAT/Service Tax/Transport charges/insurance etc.
14. Rates quoted should be valid at least for 90 days from the date of opening.
15. The rate should be quoted both in words and figure.
16. The payment will be made by account payee cheque for local supplier or by Demand Draft for outstation supplier on production of pre-receipted bill in triplicate only after satisfactory supply of the materials in good condition and as per specification.
17. TDS shall be deducted from the bill at the time of payment as per latest Govt. orders.
18. Performance Security @ the rate of 5% of the total amount will have to be deposited within 15 days from the date of issue of the supply order by the successful bidder and the same will be released after 30 days beyond the completion of warranty period. However this condition may not be applicable for KB/NCCF or similar Co-op/Govt stores.

**Executive Engineer
MID, CWC, Shillong**

Copy to:

1. The Superintending Engineer, Meghna Circle, CWC, Panchayat Road, Silchar(Assam)
2. The Director, SM Directorate, CWC, Sewa Bhawan, RK Puram , New Delhi with the request to publish at CWC Web site. A soft copy is emailed to smdte@nic.in for necessary action please.
3. The AAO, Accounts Branch, Meghna Investigation Division, Yingyar Villa, Lumpynggad, Shillong-14 with the request to present at the time of opening of quotation.
4. Notice board of office of Executive Engineer, Meghna Investigation Division, Yingyar Villa, Lumpynggad, Shillong.